



- 1) "Level In" input is 0V to +10V.
- 2) "PD In" input is:
Jumper up: amplified detector, 0V to +2.5V into 50 Ohms.
Jumper down: straight PD, 0µA to +1mA into 50 Ohms.
- 3) Audible alarm device has jumper so that other than +5V can be used.
- 4) Can add front end logarithmic amp to PD input, such as TI's LOG114.

<div style="text-align: center;">REVISIONS</div>		
University of Toronto Quantum Optics Group		ALPS - AOM Laser Power Stabilization J.Thywissen Labs
Drawn: A.Stummer June '08 Edited: 17/07/2008 17:21:54		Sheet: 1/1
		REV: A2